



PCR Quantification Detailed Report
PCR Base Line Subtracted Curve Fit Data (FAM)
Contains All Available Data

General Data

Data File Name	Luitpold Salivary DNA FIV C 5-31-11.opd
Data File Path	C:\Program Files\Bio-Rad\iQ5\Users\Craig
Collected Data	Collected Data
Current Date	5/31/2011 9:28:01 PM
Run Date	5/31/2011 6:02:50 PM
User aborted the run	No
Active RMEs	Original
Active Well Factors	Dynamic
Background Readings Valid	No, data is 341 day(s) old.
RME Valid	Yes
Well Factors Valid	No, data is 711 day(s) old.
Plate Setup File Name	Luitpold Salivary DNA FIV C 5-27-11.pts
Plate Setup File Path	C:\Program Files\Bio-Rad\iQ5\Users\Craig
Protocol File Name	FIV-PPR.C36.tmo
Protocol File Path	C:\Program Files\Bio-Rad\iQ5\Users\Jesse
Computer name	HP30948312737
Created by app	iQ5.exe (v2.0.148.60623. (OS-Microsoft Windows NT 5.1.2600.0.Service Pack 2, CLR-1.1.4322.2032, Culture-en-US).)
Created by user	BioRad\admin
Creation Date	5/31/2011 6:02:50 PM
Created in Security Edition	No
Last Creation GUID	d4a7af8b-0800-4892-9b56-ee5d5f0fd2e1
Modified by user	BioRad\admin
Last modified date	5/31/2011 6:02:50 PM
OS Build and Service Pack	2600 (Service Pack 2)
Report differs from last save	No

Notes:

Protocol:


Cycle 1: (1X)		
Step 1:	55.0 °C	for 02:00.
Cycle 2: (1X)		
Step 1:	95.0 °C	for 08:30.
Cycle 3: (45X)		
Step 1:	95.0 °C	for 00:15.
Step 2:	60.0 °C	for 01:00.
Data collection and real-time analysis enabled.		

Modified Protocol:

Protocol unchanged


PCR Quantification Data

PCR Amp/Cycle Chart



Standard Curve Data

Standard Curve Chart



Fluor	PCR	R	Slope	y-Intercept
	Efficiency(%)	Squared		
FAM	108.2	0.996	-3.140	36.814

Fluor	Units	Quantity	Original
	Changed?	Units	Units
FAM	No	copy number	copy number

Number of valid standard wells: **None**

Standard Curve Spreadsheet Data

Fluor	Well	Type	Ident.	Rep	Ct	Log SQ	SQ	SQ Mean	SQ SD	Ct Mean	Ct SD	Set Point
FAM	A01	Std	-	1	14.77	7.000	1.00E+07	1.00E+07	0.00E+00	14.70	0.102	N/A
FAM	A02	Std	-	1	14.63	7.000	1.00E+07	1.00E+07	0.00E+00	14.70	0.102	N/A
FAM	A04	Unkn	4653 d64	1	33.20	1.152	1.42E+01	3.30E+01	2.65E+01	32.31	1.246	N/A
FAM	A05	Unkn	4653 d64	1	31.43	1.714	5.17E+01	3.30E+01	2.65E+01	32.31	1.246	N/A
FAM	A07	Unkn	4666 d64	5	33.13	1.173	1.49E+01	1.67E+01	2.53E+00	32.98	0.208	N/A
FAM	A08	Unkn	4666 d64	5	32.84	1.267	1.85E+01	1.67E+01	2.53E+00	32.98	0.208	N/A
FAM	B01	Std	-	2	17.67	6.000	1.00E+06	1.00E+06	0.00E+00	17.70	0.046	N/A
FAM	B02	Std	-	2	17.73	6.000	1.00E+06	1.00E+06	0.00E+00	17.70	0.046	N/A
FAM	C01	Std	-	3	21.15	5.000	1.00E+05	1.00E+05	0.00E+00	21.19	0.064	N/A
FAM	C02	Std	-	3	21.24	5.000	1.00E+05	1.00E+05	0.00E+00	21.19	0.064	N/A
FAM	C04	Unkn	4654 d64	2	31.27	1.764	5.81E+01	7.51E+01	2.40E+01	30.96	0.444	N/A
FAM	C05	Unkn	4654 d64	2	30.65	1.964	9.21E+01	7.51E+01	2.40E+01	30.96	0.444	N/A
FAM	C07	Unkn	4667 d64	6	31.15	1.804	6.37E+01	5.05E+01	1.87E+01	31.52	0.517	N/A
FAM	C08	Unkn	4667 d64	6	31.88	1.571	3.73E+01	5.05E+01	1.87E+01	31.52	0.517	N/A
FAM	D01	Std	-	4	24.69	4.000	1.00E+04	1.00E+04	0.00E+00	24.67	0.021	N/A
FAM	D02	Std	-	4	24.66	4.000	1.00E+04	1.00E+04	0.00E+00	24.67	0.021	N/A
FAM	E01	Std	-	5	27.72	3.000	1.00E+03	1.00E+03	0.00E+00	27.73	0.007	N/A
FAM	E02	Std	-	5	27.73	3.000	1.00E+03	1.00E+03	0.00E+00	27.73	0.007	N/A
FAM	E04	Unkn	4658 d64	3	31.59	1.663	4.60E+01	4.26E+01	4.79E+00	31.70	0.153	N/A
FAM	E05	Unkn	4658 d64	3	31.81	1.594	3.92E+01	4.26E+01	4.79E+00	31.70	0.153	N/A
FAM	E07	Unkn	4669 d64	7	38.50	-0.536	2.91E-01	2.37E-01	7.57E-02	38.81	0.443	N/A
FAM	E08	Unkn	4669 d64	7	39.12	-0.736	1.84E-01	2.37E-01	7.57E-02	38.81	0.443	N/A
FAM	F01	Std	-	6	30.37	2.000	1.00E+02	1.00E+02	0.00E+00	30.52	0.214	N/A
FAM	F02	Std	-	6	30.67	2.000	1.00E+02	1.00E+02	0.00E+00	30.52	0.214	N/A
FAM	G01	Std	-	7	33.92	1.000	1.00E+01	1.00E+01	0.00E+00	33.28	0.904	N/A
FAM	G02	Std	-	7	32.64	1.000	1.00E+01	1.00E+01	0.00E+00	33.28	0.904	N/A
FAM	G04	Unkn	4663 d64	4	28.18	2.751	5.64E+02	5.56E+02	1.12E+01	28.20	0.027	N/A
FAM	G05	Unkn	4663 d64	4	28.22	2.739	5.48E+02	5.56E+02	1.12E+01	28.20	0.027	N/A
FAM	G07	Unkn	4670 d64	8	35.81	0.318	2.08E+00	2.23E+00	2.06E-01	35.72	0.127	N/A
FAM	G08	Unkn	4670 d64	8	35.64	0.375	2.37E+00	2.23E+00	2.06E-01	35.72	0.127	N/A
FAM	H01	NTC	-	1	N/A	N/A	0.00E+00	0.00E+00	0.00E+00	.00	0.000	N/A
FAM	H02	NTC	-	1	N/A	N/A	0.00E+00	0.00E+00	0.00E+00	.00	0.000	N/A

Run Parameters

Hot Start?	No
Temperature Control Mode:	Algorithmic
Volume:	25 ul

Data Analysis Parameters

Display Controls

Fluor	Display Mode
FAM	SinglePoint

Data Selection

Fluor	Data Window	Center
--------------	--------------------	---------------

FAM	Size 99%	End
-----	-------------	-----

Digital Filtering

Fluor	Global Filter Enabled?	PCR Digital Filter Type	Smoothing Filter Desired Width
FAM	Off	Weighted Mean	5

PCR Data Analysis Method

Fluor	Data Analysis Method
FAM	PCR Base Line Subtracted Curve Fit

PCR Baseline Data Analysis Parameters

Baseline Calculation

Fluor	Baseline Method	Auto Baseline Cycle Calculation?	Global Baseline Cycles	
FAM	Data Window	Yes	Start N/A	End N/A

Overriden Baseline Cycles None

Threshold Calculation

Fluor	Use Auto Threshold?	Auto Calculated Threshold Value	User Defined Threshold Value
FAM	Yes	58.17	58.17

Excluded Wells

Excluded Well Count: 0

Modified Wells

Modified Well Count: 0

End